

Date/Time: **11/11/2020**
12:52:16 PM
 Username: **ADMIN**

UDP Report

Instrument #: **CT0156(1)**

Version #: **V4.72 A**

400057C MAX

400057C MAX

Test Name	Version	Extraction Type	Master Mix Format				Ct Calculation	Lysis Heat Time
Notes	Channel Settings	475/520	530/565	585/630	630/665	680/715	Lysis Temperature	Sample Tip Height
							Sample Volume	Wash Volume
							Neutralization Volume	DNase Heat Time
							Liquid Level Sensing Enabled	
400057C MAX	1	ExK DNA-2 - 4-Snap	Type 4: Lyophilized MM with Rehydration Buffer				Call Ct at Threshold Crossing	9
--	Alias	--	HSV2	HSV1	VZV	SPC	62	1600
--	Gain	--	60	60	60	60	937.5	500
--	Threshold	--	200	200	200	200	--	--
--	Excitation Channel 475/520	--	0	0	0	0		
--	Excitation Channel 530/565	0	--	1	0	0		
--	Excitation Channel 585/630	0	0	--	3.5	0		
--	Excitation Channel 630/665	0	0	5	--	0		
--	Excitation Channel 680/715	0	0	0	1	--		

Disabled

Step Name: New Step
Profile Type: Hold **Cycles:** 1

Type	Time (s)	Temp (°C)	Detect
--	300	99	No

Step Name: New Step
Profile Type: 3 - Temperature **Cycles:** 40

Type	Time (s)	Temp (°C)	Detect
--	11.5	99	No
--	32.5	58	Yes
--	20	72	No

	HSV2			
HSV2 - PCR	✓	✓	✗	✗
SPC - PCR	✓	✗	✓	✗
Results	➕ POS	➕ POS	➖ NEG	ⓘ UNR

	HSV1			
HSV1 - PCR	✓	✓	✗	✗
SPC - PCR	✓	✗	✓	✗
Results	➕ POS	➕ POS	➖ NEG	ⓘ UNR

	VZV			
VZV - PCR	✓	✓	✗	✗
SPC - PCR	✓	✗	✓	✗
Results	➕ POS	➕ POS	➖ NEG	ⓘ UNR



Performed By: _____ Date: _____
 Reviewed By: _____ Date: _____

Printed: 2020-11-11 12:52:24

Page: 1/1