

Date/Time: **11/3/2021 10:41:02 PM**  
 Username: **ADMIN**

# UDP Report

Instrument #: **MX0000(1)**  
 Version #: **V4.72 A**

## PanAsp350058CMAX

Test Name	Version	Extraction Type	Master Mix Format				Ct Calculation		Lysis Heat Time
			475/520	530/565	585/630	630/665	680/715	Lysis Temperature	Sample Tip Height
Notes	Channel Settings						Sample Volume	Wash Volume	
							Neutralization Volume	DNase Heat Time	
							Liquid Level Sensing Enabled		
PanAsp350058CMAX	3	ExK DNA-1 (Urine) - 4-Snap	Type 4: Lyophilized MM with Rehydration Buffer				Call Ct at Threshold Crossing		9
--		Alias	--	--	Pan Asp	--	SPC	62	1600
		Gain	--	--	60	--	60	500	500
		Threshold	--	--	400	--	200	--	--
		Ct Min	--	--	--	--	--	--	--
		Ct Max	--	--	32	--	--	--	--

Disabled

**Step Name:** New Step  
**Profile Type:** Hold **Cycles:** 1

Type	Time (s)	Temp (°C)	Detect
--	300	99	No

**Step Name:** New Step  
**Profile Type:** 2 - Temperature **Cycles:** 40

Type	Time (s)	Temp (°C)	Detect
--	14	99	No
--	46.6	64	Yes



Performed By: \_\_\_\_\_ Date: \_\_\_\_\_  
 Reviewed By: \_\_\_\_\_ Date: \_\_\_\_\_

Printed: 2021-11-03 22:41:03

Page: 1/1