

Date/Time: 5/22/2020 8:16:18 AM
 Username: FSVC

UDP Report

Instrument #: CT1557(1)
 Version #: V4.72 A

Atyp-350003AMAXnas

Test Name	Version	Extraction Type	Master Mix Format				Ct Calculation	Lysis Heat Time	
Notes	Channel Settings	475/520	530/565	585/630	630/665	680/715	Lysis Temperature	Sample Tip Height	
							Sample Volume	Wash Volume	
							Neutralization Volume	DNase Heat Time	
							Liquid Level Sensing Enabled		
Atyp-350003AMAXnas	7	ExK DNA-1 (Urine) - 4-Snap	Type 4: Lyophilized MM with Rehydration Buffer				Call Ct at Threshold Crossing	30	
--		Alias	Leg spp	Mpneumo	--	Cpneumo	SPC	62	1600
		Gain	60	60	--	60	60	500	500
		Threshold	100	100	--	100	100	--	--
							Disabled		

Step Name: New Step
Profile Type: Hold **Cycles:** 1

Type	Time (s)	Temp (°C)	Detect
--	120	99	No

Step Name: New Step
Profile Type: 2 - Temperature **Cycles:** 45

Type	Time (s)	Temp (°C)	Detect
--	14.4	99	No
--	44.4	62	Yes



Performed By: _____

Date: _____

Printed: 2020-05-22 08:16:19

Reviewed By: _____

Date: _____

Page: 1/1