

Date/Time: 5/22/2020 4:16:12 AM
 Username: FSVC

UDP Report

Instrument #: CT1557(1)
 Version #: V4.72 A

Cauris-350070MAX

Test Name	Version	Extraction Type	Master Mix Format				Ct Calculation		Lysis Heat Time		
			Notes	Channel Settings	475/520	530/565	585/630	630/665	680/715	Lysis Temperature	Sample Tip Height
										Sample Volume	Wash Volume
										Neutralization Volume	DNase Heat Time
										Liquid Level Sensing Enabled	
Cauris-350070MAX	1	ExK DNA-3 - 4-Snap	Type 4: Lyophilized MM with Rehydration Buffer				Call Ct at Threshold Crossing		9		
--		Alias	--	--	C auris	--	SPC	62	1600		
--		Gain	--	--	60	--	60	700	500		
--		Threshold	--	--	200	--	200	--	--		
								Disabled			

Step Name: New Step
Profile Type: Hold **Cycles:** 1

Type	Time (s)	Temp (°C)	Detect
--	300	99	No

Step Name: New Step
Profile Type: 3 - Temperature **Cycles:** 40

Type	Time (s)	Temp (°C)	Detect
--	10	99	No
--	29.5	58	Yes
--	14.5	70	No



Performed By: _____

Date: _____

Printed: 2020-05-22 04:16:13

Reviewed By: _____

Date: _____

Page: 1/1